

Controlling the quality factor of a tuning-fork resonance between 9 and 300 K for scanning-probe microscopy

G Ctistis, E H Frater, S R Huisman, J P Korterik, J L Herek, W L Vos and P W H Pinkse

MESA+ Institute for Nanotechnology, University of Twente, 7500 AE Enschede, The Netherlands

E-mail: g.ctistis@utwente.nl and P.W.H.Pinkse@utwente.nl

Received 3 May 2011, in final form 20 July 2011

Published 26 August 2011

Online at stacks.iop.org/JPhysD/44/375502

Abstract

We study the dynamic response of a mechanical quartz tuning fork in the temperature range from 9 to 300 K. Since the quality factor Q of the resonance strongly depends on temperature, we implement a procedure to control the quality factor of the resonance. We show that we are able to dynamically change the quality factor and keep it constant over the whole temperature range. This procedure is suitable for applications in scanning-probe microscopy.

(Some figures in this article are in colour only in the electronic version)

1. Introduction

Quartz crystal oscillators are widely used as a frequency reference. The mechanical resonance is intrinsically stable and the crystal's piezoelectric effect allows for direct electronic interfacing. Quartz tuning forks have proven to be extremely useful as sensors for temperature [1, 2], mass [3], pressure [4–7] and friction [8–17]. Due to their high mechanical quality factor Q between 10^3 and 10^5 they are very sensitive to environmental changes. Sub-pN forces dragging on their prongs can easily be detected. The dragging forces change the oscillating properties of the tuning fork through a change in the spring constant. This gives rise to a shift in the resonant frequency and a corresponding change in the amplitude of the vibrating tuning fork. This change in the vibration amplitude changes the current through the tuning fork, which is generated through the piezoelectric effect. Thus, by measuring the changes of the electric signal, which can be measured very accurately, one can diagnose the prevailing forces.

This sensitivity has made quartz tuning forks become popular in two fields of research: temperature measurement of liquid He, where the resonance frequency is a measure for the temperature of the surrounding liquid [18–20], and in scanning-probe microscopy (SPM), where the forces between tip and sample are used to image surfaces with atomic resolution [8–10, 21]. More and more low-temperature scanning-probe microscopes are used nowadays [22–24]

where the height of the tip is controlled via the shear forces of a vibrating tip attached to a tuning fork. An accurate control of the resonance response of the tuning fork is therefore highly desired because the performance of the height control critically depends on the resonance properties of the tuning fork. These properties are thereby described by the quality factor Q of the resonance. It is defined as 2π times the mean stored energy divided by the work per cycle or equivalently by $Q = \omega_0/\Delta\omega$, where ω_0 is the resonance frequency and $\Delta\omega$ the width of the resonance. A high quality factor Q normally results in an undesirably slow response of the system [25] since the response time τ of the system is related to the quality factor by $\tau = 2Q/\omega_0$. Controlling the quality factor of the resonance provides a solution to this problem. It has already been shown [25–31] that quality-factor control is possible. Yet, up to now, it has only been applied to measurements in liquids, in ultrahigh vacuum, or at liquid He temperatures, but it has not been used to measure temperature-dependent system dynamics.

Here, we present the dynamic control of the quality factor of the resonance of a tuning fork over a large range of temperatures. We are able to adjust the quality factor of the resonance to a desired value and keep it constant while changing the temperature, which is vital for applications purposes in near-field scanning optical microscopy (NSOM).

The outline of the paper is as follows: a brief introduction into the oscillator tuning-fork system and its temperature dependence is followed by a description of the quality-factor




Figure 1. Schematic of the electric circuit for the tuning-fork excitation. The parallel amplifier and capacitance are for compensation of the parasitic capacitance of the tuning fork. The dashed box shows the part responsible for the quality-factor control, consisting of a 90° phase shifter (Q -control) and an amplifier (gain).

controlled system and the presentation of the data showing that we are able to control the quality-factor of the resonance over a wide range of temperatures. A short conclusion with prospects for possible experiments closes the paper.

2. The tuning-fork system

The experiments were performed both under ambient conditions as well as in a cryostat at a temperature between 9 and 300 K and a base pressure of less than 1×10^{-4} mbar. Figure 1 shows a schematic of the electric circuit to measure the electric response of the tuning fork. A sinusoidal signal from an external function generator (IN) drives the tuning fork at a constant frequency with a constant amplitude. The resulting oscillation amplitude and phase shift are detected by a lock-in amplifier (OUT). The constant frequency of the function generator acts also as a reference for the lock-in. The additional feedback circuit (grey dashed box) consisting of an amplifier and a phase shifter (e.g. a fixed phase of 90°) is responsible for the quality-factor control and works as described, for example, in [26]. The signal of this feedback circuit is amplified (gain) and used to excite the tuning fork in addition to the signal from the function generator (Σ). The phase shifter controls thereby the time delay (phase) between the excitation and the tuning-fork signal, which results in a damping or amplification of the tuning-fork excitation. If the driving signal and the control signal have a phase difference of $\pi/2$ we can cancel the natural damping of the tuning fork. By adjusting the gain factor g and the phase (corresponding time delay t_0) we can increase or decrease the quality factor Q_{eff} [29, 31]:

$$Q_{\text{eff}}(g, t_0) = \frac{1}{1/Q - g \sin(\omega_0 t_0)}. \quad (1)$$

The oscillator system itself consists of a standard quartz tuning fork with an eigenfrequency of $\omega_0 \approx 2\pi \times 2^{15}$ Hz. One prong is $l_r = 3.9$ mm long, $t_r = 600 \mu\text{m}$ thick and $w_r = 330 \mu\text{m}$ wide. Since the tuning fork is excited electrically, we expect the so-called anti-phase mode to be the dominant oscillation mode [32–34]. In part of the experiments




Figure 2. Amplitude (upper panel) and phase (lower panel) response of a bare tuning fork (circles) and a tuning fork with a glued fibre (squares) at room temperature. Additionally, fitted curves (dashed lines) determining the quality factors of the two tuning forks are shown. The black dotted lines indicate the centre frequency.

an optical fibre was glued to one prong of the tuning fork with epoxy glue for the use in temperature-dependent NSOM measurements. This changes not only the thickness of one prong by approximately half of the cross-sectional area of the fibre ($t = 617 \mu\text{m}$) but also changes the mass of this prong compared with the second unloaded prong. The tuning fork is therefore not balanced anymore but asymmetric. One expects changes in the response of the tuning fork as compared with a bare one as has already been reported before [32–34]. The elastic constant of quartz and its density are $E = 8.68 \times 10^{10} \text{ N m}^{-2}$ and $\rho = 2649 \text{ kg m}^{-3}$, respectively [35]. The resonance frequency of the tuning-fork-fibre system is calculated using [36]

$$\nu_0 = \sqrt{\frac{E}{4} \frac{t_r (w_r/l_r)^3}{\alpha \rho l_r t_r w_r}} = 33\,690 \text{ Hz}. \quad (2)$$

The measured eigenfrequency of the tuning-fork-fibre system is $\nu_{\text{meas}} = 33\,550 \text{ Hz}$. The calculation matches the experimentally determined frequency very well given the slight difference in values due to the evaporated contacts, the glue, and the attached fibre.

Figure 2 shows the amplitude (upper panel) and the phase (lower panel) response of a bare tuning-fork (black circles) as well as of a tuning fork-fibre system (green squares). Both resonance curves were taken under ambient conditions. It is clearly visible that gluing a fibre to one prong of the tuning fork changes the oscillator drastically. The resonance blue-shifts by 1 kHz and its amplitude is damped. The dashed lines

in figure 2 show the resonances' amplitude and phase response, which is calculated with

$$A(\omega) = \frac{A_0}{m} \frac{1}{\sqrt{(\omega_0^2 - \omega^2)^2 + \gamma^2 \omega^2}} \quad (3)$$

$$\phi(\omega) = \arctan\left(\frac{\gamma \omega}{\omega_0^2 - \omega^2}\right), \quad (4)$$

where ω_0 is the resonance frequency, m is the mass of the system, A_0 the driving amplitude and γ the damping. The model matches the experiment well and reveals quality factors $Q = \omega_0/\Delta\omega$ of 5400 for the bare tuning fork and 500 for the tuning-fork-fibre system, respectively.

In the next step we determined the temperature dependence of the resonance response of the bare tuning fork and the tuning-fork-fibre system. We transferred the oscillator system into a small cryophysics cryostat with a model 22CTI cryodyne closed-cycle helium cooler, which allowed us to measure at temperatures between 9 and 300 K. Figure 3 shows the acquired data in a colour-scale representation where the colours represent the amplitude of the resonance measured versus the frequency for each temperature. The temperature-dependent response of the bare tuning fork (lower curve) shows two major features. First of all only a small shift of the resonance frequency is measured throughout the whole temperature regime. Down to a temperature of about 70 K this frequency shift is well described by

$$\omega = \omega_0[1 - \alpha \cdot (T - T_0)^2], \quad (5)$$

with $\alpha = 3 \times 10^{-8} \text{ K}^{-2}$ and $T_0 = 300 \text{ K}$. The value for α corresponds thereby with the value given by the manufacturer [37] and is inherent to the quartz crystal and to the orientation in which it is cut. Below 77 K the resonance shift is flattened, which is mainly due to the change in the elastic constants of the quartz [38]. Second, the resonance stays very sharp throughout the whole temperature range and has a quality factor of $Q \approx 14000$ (see also figure 4). We can assume that the measured temperature-dependent response of the bare tuning fork is universal due to the manufacturer's design. The effect of the vacuum in the cryostat on the resonance response can be neglected in our experiments because it leads to a relative shift of the resonance in the order of 10^{-4} .

In contrast to the bare tuning fork, the temperature dependence of the resonance response of the combined tuning-fork-fibre system is more complex (figure 3, upper curve). This is caused by the glue and the attached fibre, which lead to a lower quality factor compared with the bare tuning fork. Overall, a blue-shift of the resonance frequency can be observed, which corresponds well with an increased stiffness of the system. More evident is that the width of the resonance decreases with decreasing temperature. This means that at low temperatures the oscillator system has a higher quality factor than at room temperature. Another feature that is apparent in the resonance response is at $T \approx 60 \text{ K}$. One can clearly observe a splitting and anti-crossing behaviour of the resonance where the two branches interchange oscillator strength, i.e. amplitude. Whereas the dominant mode is the fundamental




Figure 3. Measurement of the amplitude of the oscillation of a bare tuning fork (lower curve) and of a tuning fork with an attached fibre (upper curve) versus frequency and temperature. The change in the resonance frequency and width is significant in the second case.

anti-symmetric mode of the tuning fork, the second mode cannot be the fundamental symmetric mode because the symmetric and anti-symmetric modes are at least 10 kHz apart [33]. Even considering the asymmetry of the tuning fork because of the loaded prong using a two-coupled-oscillator model [32–34] cannot fully explain the observed resonance splitting. An explanation might be that the oscillation mode of the loaded prong alone changes with temperature due to the changing elastic properties of the glue that was used to attach the fibre. This assumption is strengthened by the fact that the observed splitting strongly depends on the amount of glue used, the mass of the fibre attached, and how the fibre is attached to one prong of the tuning fork, thus showing a case-by-case response. A detailed analysis of the response is beyond the scope of this paper.

3. Controlling the quality factor

It is evident from these measurements that in order to perform shear-force controlled scanning-probe measurements, we need to control the quality factor of the resonance and keep it stable in the complete temperature range. As a rule of thumb the quality factor of a tuning-fork system used in SPM should in general lie between 300 and 1000. In this range the feedback system for the shear-force measurements is found to be well balanced, allowing for stable measurements. From our data shown in figure 3 we can extract the corresponding quality factors for our systems from the resonance width and the resonance frequency of the tuning-fork response. Throughout the whole temperature range these values are shown in figure 4 as open symbols, squares for the bare tuning fork and circles for the tuning-fork-fibre system. It is seen that the bare tuning fork has quality factors exceeding $Q = 12000$. More striking is the dependence for the tuning-fork-fibre system. It is obvious that in the high-temperature region the quality factor is in the desired range ($Q \approx 500$). Yet, by decreasing temperature the quality factor increases by a factor of ~ 5 to $Q = 2300$




Figure 4. Measured quality factor versus temperature for a bare tuning fork (open squares) and a tuning fork with a glued fibre (open circles). The shaded region indicates the desired range of quality factors where shear-force measurements are possible. Applying quality-factor control results in a near-constant quality factor for both systems (filled symbols).

at 9 K. Thereby the change in quality factor is not monotonic but changes also due to the observed mode anti-crossing.

The filled symbols in figure 4 display our results after applying our electronic phase compensation, i.e. our quality-factor control. We succeeded in reducing the quality factor of both the bare tuning fork and the tuning-fork-fibre system and kept it nearly constant throughout the whole temperature range.

To test our system’s dynamic performance we continuously changed the gain settings for the quality-factor control. Figure 5 shows the resonance response of a tuning-fork-fibre system at a temperature of 9 K. We recorded the amplitude (upper panel) as well as phase (lower panel) versus the frequency. The starting value of $Q = 2300$ (compare also figure 4) is achieved without any quality-factor control. As is evident from the graph, we can change the quality factor of the tuning fork smoothly down to $Q = 40$ by changing the gain in the phase shifter of the quality-factor-control unit (figure 1, grey box). The good performance of our quality-factor control is also apparent in the recorded phase, a smooth transition from high- to low- Q is visible.

The only restricting factor still prominent in the measurements is the noise level. For lower quality factors the signal-to-noise level is the limiting parameter to resolve the resonance of the tuning fork clearly. If the main noise source is the electronics that were used, this restriction can be reduced significantly. Applying an electronic compensation scheme where the reference signal and the signal from the tuning fork are subtracted, we achieved reduction of the noise level by at least a factor of 10 (data not shown here), thereby increasing the performance of the system for the entire range of quality factors.




Figure 5. Quality-factor control measurements of the resonance at $T = 9$ K for a tuning fork with an attached fibre. Upper panel: amplitude response of the tuning fork. The quality factor of the resonance was tuned between $Q = 40$ and 2300. Lower panel: simultaneously recorded phase of the oscillator.




Figure 6. Shear-force topography scan of a test grating at room temperature. For this scan an adjusted quality factor of $Q \approx 350$ was used. Between the left and right parts of the image the gain in the height feedback control was changed. Right panel: the cross section was taken along the dashed line in the topography image.

With the uncontrolled high value for Q we could not achieve any shear-force controlled measurement of a topography. Yet, by changing the quality factor and keeping it constant ($Q \approx 350$) we are now able to measure the topography of a standard test grating [39] with our tuning-fork-fibre system, as shown in figure 6 (left panel). The displayed squares have a height of 20 nm, a pitch size of $1.5 \mu\text{m}$ and a periodicity of $3 \mu\text{m}$. There is visible difference between the left and right part of the recorded image visible. The left part has more noise along the fast scan direction of the microscope. Changing the gain of the height feedback halfway during the scan led to reduced noise and a higher contrast in the topography. This is visible in more detail in the

cross-sectional scan (figure 6, right panel) along the dashed line of the topography image. The height of the pitches deduced from this scan matches the height given by the manufacturer and shows the good performance of the shear-force quality-factor controlled scanning system.

4. Conclusions and outlook

We presented our results on the dynamic control of the quality factor of a quartz tuning fork in a large temperature range. We showed that we are able to compensate for the temperature-dependent change in the resonance response of a tuning fork. Moreover, we applied our compensation to a tuning-fork-fibre system and achieved keeping the quality factor in a regime where shear-force measurements are possible throughout the whole temperature range. We are also able to smoothly tune the quality factor at each temperature to a certain value. To be able to measure topography with the same quality factor of the tuning fork is a major step to a more efficient and straightforward use of such scanning-probe microscopy systems and allows their use over the complete temperature range as well as in all sorts of environments.

As a result of our experiments, dynamic temperature-dependent near-field optical measurements have become possible. The method can be used, e.g. to measure locally resolved structural phase transitions, such as domain formation, by optical means with varying temperature and thus deduce the optical properties of sub-wavelength structures near a transition point. Furthermore, the temperature dependence of light propagation in nanophotonic objects can be studied.

Acknowledgments

We would like to thank Frans Segerink, Allard Mosk, Herman Offerhaus, Cock Harteveld and Dick Veldhuis for discussions and help with setting up equipment. PP acknowledges financial support by FOM Projectruimte. GC acknowledges financial support by NWO-Nano. This work is also part of the research program of FOM, which is financially supported by NWO.

References

- [1] Smith W L and Spencer W J 1963 *Rev. Sci. Instrum.* **34** 268
- [2] Jun X, Bo Y, Xin L and Juan C 2007 *Phys. Scr.* **T129** 316
- [3] Sauerbrey G 1959 *Z. Phys.* **155** 206
- [4] Christen M 1983 *Sensors Actuators* **4** 555
- [5] Kokubun K, Hirata M, Ono M, Murakami H and Toda Y 1985 *J. Vac. Sci. Technol. A* **3** 2184
- [6] Kokubun K, Hirata M, Ono M, Murakami H and Toda Y 1987 *J. Vac. Sci. Technol. A* **5** 2450
- [7] Stemme G 1991 *J. Micromech. Microeng.* **1** 113
- [8] Toledo-Crow R, Yang P C, Chen Y and Vaez-Iravani M 1992 *Appl. Phys. Lett.* **60** 2857
- [9] Betzig E, Finn P L and Weiner J S 1992 *Appl. Phys. Lett.* **60** 2484
- [10] Karrai K and Grober R D 1995 *Appl. Phys. Lett.* **66** 1842
- [11] Pendry J B 1997 *J. Phys.: Condens. Matter* **9** 10301
- [12] Grober R D, Acimovic J, Schuck J, Hessman D, Kindlemann P J, Hespanha J, Morse A S, Karrai K, Tiemann I and Manus S 2000 *Rev. Sci. Instrum.* **71** 2776
- [13] Wei P K and Fann W S 2000 *J. Appl. Phys.* **87** 2561
- [14] Karrai K and Tiemann I 2000 *Phys. Rev. B* **62** 13174
- [15] Leuschner M, Schüttler M and Giessen H 2001 *J. Microsc.* **202** 176
- [16] Stipe B C, Mamin H J, Stowe T D, Kenny T W and Rugar D 2001 *Phys. Rev. Lett.* **87** 096801
- [17] Hoppe S, Ctistis G, Paggel J J, and Fumagalli P 2005 *Ultramicroscopy* **102** 221
- [18] Blaauwgeers R *et al* 2007 *J. Low. Temp. Phys.* **146** 537
- [19] Skyba P 2010 *J. Low. Temp. Phys.* **160** 219
- [20] Pentti E M, Tuoriniemi J T, Salmela A J and Sebedash A P 2008 *J. Low. Temp. Phys.* **150** 555
- [21] Giessibl F J 2003 *Rev. Mod. Phys.* **75** 949
- [22] Rychen J, Ihn T, Studerus P, Herrmann A, Ensslin K, Hug H J, van Schendel P J A and Güntherodt H J 2000 *Rev. Sci. Instrum.* **71** 1695
- [23] Wintjes N, Lange M, van Vörden D, Karacuban H, Utzat D and Möller R 2010 *J. Vac. Sci. Technol. B* **28** C4E21
- [24] Saitoh K, Hayashi K, Shibayama Y and Shirahama K 2008 *J. Low. Temp. Phys.* **150** 561
- [25] Albrecht T R, Grütter P, Horne D and Rugar D 1991 *J. Appl. Phys.* **69** 668
- [26] Anczykowski B, Cleveland J P, Krüger D, Elings V and Fuchs H 1998 *Appl. Phys. A* **66** S885
- [27] Tamayo J, Humphris A D L and Miles M J 2000 *Appl. Phys. Lett.* **77** 582
- [28] Tamayo J, Humphris A D L, Owen R J and Miles M J 2001 *Biophys. J.* **81** 526
- [29] Rodriguez T R and Garcia R 2003 *Appl. Phys. Lett.* **82** 4821
- [30] Morville J, Liu J, Callegari A and Chergui M 2005 *Appl. Phys. Lett.* **86** 064103
- [31] Jahng J, Lee M, Noh H, Seo Y and Jhe W 2007 *Appl. Phys. Lett.* **91** 023103
- [32] Naber A 1999 *J. Microsc.* **194** 307
- [33] Castellanos-Gomez A, Agrait N and Rubio-Bollinger G 2009 *Nanotechnology* **20** 215502
- [34] Castellanos-Gomez A, Agrait N, and Rubio-Bollinger G 2009 *Ultramicroscopy* **111** 186
- [35] Lide D R (ed) 2000 *CRC Handbook of Chemistry and Physics* (Boca Raton, FL: CRC Press)
- [36] Sarid D 1991 *Scanning Force Microscopy* ed M Lapp and H Stark (Oxford: Oxford University Press)
- [37] Micro Crystal AG <http://www.microcrystal.com>
- [38] Fukuhara M, Sanpei A and Shibuki K 1997 *J. Mater. Sci.* **32** 1207
- [39] NT-MDT Co, TGQ1 test grating, <http://www.ntmdt-tips.com>